



CERTIFICADO DE ANÁLISE

PRODUTO:	MICROPIPETA DENUDACAO INGAMED 150UM CX 20UN	INSPEÇÃO:	006923
APRESENTAÇÃO:	CAIXA COM 20 UNIDADES	LOTE:	24050336
DT. PRODUÇÃO:	30/04/2024	DT. VALIDADE:	31/03/2027

USO INDICADO	Indicada para a manipulação de oócitos e embriões.
APLICAÇÃO	Usada na manipulação de oócitos e embriões, evitando arranhões na superfície ou rupturas.
ARMAZENAMENTO	Armazenar em temperatura ambiente, protegido do calor e umidade.
MATERIA PRIMA E INSUMOS	Policarbonato

ANÁLISES E RESULTADOS		ESPECIFICAÇÃO	RESULTADO
BIOMEA	BIOENSAIO COM EMBRIÕES DE CAMUNDONGO	TAXA DE BLASTOCISTOS = 80%	95%
ESTERIL RG	ESTERILIDADE	AUSÊNCIA DE MICRO ORGANISMOS VIÁVEIS	CONFORME
DIRETIVA EEC 93/42	DIRETIVA EEC 93/42	CONFORMIDADE COM A DIRETIVA EEC/93/42	CONFORME

RESPONSÁVEL ANÁLISE:	ELIZIA
RESPONSÁVEL CONFERÊNCIA:	ELIZIA



Reproline medical GmbH
 Zeissstr. 15
 53359 Rheinbach
 Germany



140 Hale Street
 Haverhill, MA 01830
qc@embryotech.com

ELI Accession Number: REP-4094-0424

Date of completion: 04-30-2024

Lot number: 442402/14/01

Reference number: 447150/20

Description of test article(s): Soft Denudation Tip - 150µm

Assay system requested by customer: 1 mL of culture medium was placed in a tube with the test articles (3) for 30-minutes at 37°C and 5% CO₂. Post incubation three 12.5µl drops of the culture medium was extracted from the test article tube and placed in the corresponding wells of a culture dish; 7 one-cell mouse embryos were added to each of the three wells and cultured for 96-hours.

Control assay method and results: 21 one-cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium:

20 / 21 (95 %)	1-cell to 2-cell within 24 hr
19 / 21 (90 %)	1-cell to expanded blastocyst within 96 hr

For a valid assay, *Embryotech™* requires at least 80% of one-cell stage control embryos to develop to expanded blastocyst within 96-hours.

Test assay method and results: 21 one-cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium that was extracted from the test article tube:

21 / 21 (100 %)	1-cell to 2-cell within 24 hr
20 / 21 (95 %)	1-cell to expanded blastocyst within 96 hr

Pass/Fail = Pass

Summary of observations: All test and control embryos were selected randomly from a common pool of freshly collected embryos and were cultured in the same incubator at 37°C and 5.0% CO₂. 90 percent of the control embryos developed to the expanded blastocyst stage within 96-hours. 95 percent of the test embryos cultured in the extracted culture medium developed to the expanded blastocyst stage within 96-hours.

Emm Luett

Signature
 Study Director

4/30/24

Date

Adl Gio

Signature
 Quality Reviewer

4/30/24

Date

to order of BRAZIL

Certificate of Sterilization

relating the products manufactured by us and sold mentioned in
INVOICE No. 402413720 dated May 06th, 2024

REF 447135/20	Soft Denudation Tip 135 µm	LOT 442401/03/03	MFD 01/2024	Exp. 12/2026
REF 447135/20	Soft Denudation Tip 135 µm	LOT 442401/11/03	MFD 03/2024	Exp. 02/2027
REF 447150/20	Soft Denudation Tip 150 µm	LOT 442401/02/03	MFD 01/2024	Exp. 12/2026
REF 447150/20	Soft Denudation Tip 150 µm	LOT 442402/14/01	MFD 04/2024	Exp. 03/2027
REF 447175/20	Soft Denudation Tip 175 µm	LOT 442401/12/01	MFD 03/2024	Exp. 02/2027
REF 447200/20	Soft Denudation Tip 200 µm	LOT 442401/13/03	MFD 03/2024	Exp. 02/2027
REF 447275/20	Soft Denudation Tip 275 µm	LOT 442401/11/01	MFD 03/2024	Exp. 02/2027
REF 447275/20	Soft Denudation Tip 275 µm	LOT 442401/12/02	MFD 03/2024	Exp. 02/2027
REF 447300/20	Soft Denudation Tip 300 µm	LOT 442401/11/04	MFD 03/2024	Exp. 02/2027

All raw materials were adapted optimally to the validated sterilization by gamma irradiation during the developmental phase.

We guarantee the sterility of our products according to DIN EN 11137

Reproline medical GmbH

Zeissstr. 15
53359 Rheinbach
Germany

phone: +49(0)2226/90050
fax: +49(0)2226/900555

Ludger Hoppe

Rheinbach/Germany
May 06th, 2024

DECLARATION OF CONFORMITY

We
Reproline medical GmbH
Zeissstrasse 15
53359 Rheinbach
Germany

hereby declare that products stated below

REF 447135/20	Soft Denudation Tip 135 µm	LOT 442401/03/03	MFD 01/2024	Exp. 12/2026
REF 447135/20	Soft Denudation Tip 135 µm	LOT 442401/11/03	MFD 03/2024	Exp. 02/2027
REF 447150/20	Soft Denudation Tip 150 µm	LOT 442401/02/03	MFD 01/2024	Exp. 12/2026
REF 447150/20	Soft Denudation Tip 150 µm	LOT 442402/14/01	MFD 04/2024	Exp. 03/2027
REF 447175/20	Soft Denudation Tip 175 µm	LOT 442401/12/01	MFD 03/2024	Exp. 02/2027
REF 447200/20	Soft Denudation Tip 200 µm	LOT 442401/13/03	MFD 03/2024	Exp. 02/2027
REF 447275/20	Soft Denudation Tip 275 µm	LOT 442401/11/01	MFD 03/2024	Exp. 02/2027
REF 447275/20	Soft Denudation Tip 275 µm	LOT 442401/12/02	MFD 03/2024	Exp. 02/2027
REF 447300/20	Soft Denudation Tip 300 µm	LOT 442401/11/04	MFD 03/2024	Exp. 02/2027

Classification IIa

in compliance with annex V and Annex VII

meet the provision of the
 medical Device Directive 93/42/EEC dated June 14th, 1993

The Validity of this declaration is limited to 2 Years after signing

Reproline medical GmbH

Zeissstr. 15

53359 Rheinbach
 Germany

phone: +49(0)2226/90050

fax: +49(0)2226/900555

Ludger Hoppe

Rheinbach/Germany

May 06th, 2024